



The driver installation package *PC-USB_IMB-MDU_inst.exe* installs the PC-usb / IMB-mdu driver. This driver is required to operate the following devices :


- PC-usb (F601 020)
- IMB-mdu (F120 110)

The driver installation package is located on the CD-ROMs ,*IBR Support Software* and ,*Software for Metrology and SPC* in the directory <CDROM>:\USB-Driver\PC-USB_IMB-MDU\PC-USB_IMB-MDU_inst.exe.

The most recent version can be downloaded at http://www.ibr.com/download/PC-USB_IMB-MDU_inst.zip .


1.

When the driver installation package is started, the installer asks for a directory where the driver files should be copied to :



2.

By clicking on Next, the installer copies the required files to the directory selected and automatically opens the Device Driver Installation Wizard :



3.

